



Keysight Technologies Malaysia Sdn Bhd (463532-M)
 Bayan Lepas Free Industrial Zone
 11900 Penang,
 Malaysia.



5962-0476

Certificate Of Calibration

Certificate No: 85054DMY5340036320180418

Manufacturer: Keysight Technologies

Model No: 85054D

Options Installed With Specifications: N/A

Description: Type-N Economy Calibration Kit

Serial No: MY53400363

Date of Calibration: 18-APR-2018

Temperature: (23 ± 0.5) °C

Humidity: (20 to 55) % RH

Procedure: MTA-T0107, MTA-T0126, MTA-T0127, MTA-T0128

This certifies that the equipment has been calibrated using applicable Keysight Technologies procedures in compliance with a quality management system registered to ISO 9001:2015.

As Received Conditions: Factory tested. No incoming data available.

As Shipped Conditions: At the completion of the calibration, measured values were IN SPECIFICATION at the points tested.

Remarks or special requirements:

Notes:

- 1) This calibration report shall not be reproduced, except in full.


Traceability Information: Measurements are traceable to the International System of Units (SI) via national metrology institutes (www.keysight.com/find/NMI) that are signatories to the CIPM Mutual Recognition Arrangement.

Calibration Equipment Used:

Date used: Date equipment used in this calibration.

Model Number	Model Description	Trace Number	Date Used	Cal Due Date
E8364C	PNA Network Analyzer	PD7522	18-APR-2018	30-JUL-2018
ET33717	Type-N50 Calibration Kit	PA7129	27-JUN-2018	27-MAR-2019
ET33717	Type-N50 Calibration Kit	PC7048	18-APR-2018	04-JUL-2018
ET33717,005	Type-N50 Adapter Test Kit	PA7130	27-JUN-2018	17-OCT-2018
ET33733	APC7 Calibration Kit	PC7049	27-JUN-2018	28-JUL-2018
E8364C	PNA Network Analyzer	PD7534	27-JUN-2018	11-JUN-2019

Print Date: 03-JUL-2018


 Tay Eng Su
 Quality Manager

Keysight Technologies				
	DD	MM	YY	BY:
CAL	18	04	18	ZI
DUE				