

# Failed Calpration Orthot SHIP!!!

Part No: 239105-SP

Range: 0 to 30 IN WC

Output: 0 to 10 VDC

Serial No: 4995086 Work Order: 34814

Date: 01/05/2021

Tech.: RR 11Model: 239

Supply: 24VDC

#### CALIBRATION DATA

APPLIED PRESSURE (IN WC)	TRANSDUCER OUTPUT (VDC)	NONLINEARITY ERRORS (% FS)	EXTRAPOLATED ERRORS (% FS)
0.2990	0.0857	0.050	Zero -0.147
3.2994	1.0885	0.001	
6.2897	2.0897	-0.029	Span 0.749
9.2958	3.0972	-0.050	
12.2992	4.1066	-0.041	
15.2930	5.1131	-0.031	
18.2977	6.1249	-0.003	
21.2945	7.1335	0.018	
24.2984	8.1437	0.032	
27.2923	9.1503	0.044	
30.0713	10.0842	0.050	

REJECT CALIBRATION: NONLINEARITY OUT OF SPEC.

REJECT CALIBRATION: SPAN OUT OF SPEC.

## Failed Cal. DO NOT SHIP!!!

### SPECIFICATIONS

- 1. Nonlinearity: +/- 0.045 %FS, BEST FIT STRAIGHT LINE method, ISA.#S-37.1
- 2. Zero pressure output: 0 VDC +/- 0.4 %FS
- 3. Full Scale output: 10 VDC +/- 0.4 %FS
- 4. This unit meets the specifications defined above.

### NOTES

- 1. All errors are expressed as: Percent Full-Scale output.
- 2. Consult specification sheet for additional specifications.
- 3. This calibration is certified per N.I.S.T. traceable primary standards. NIST#  $Fluke_Rpt#1500266324$

Transfer standard: 23912D30.28, Location of cal.: PCTS01

- 4. This part uses spec. record number: 239105-12-604-702-710.1
- 5. This certificate cannot be reproduced except in full, without the writter approval of Setra Systems, Inc.

#### COMMENTS

AS RECEIVED.

OPTIONS:901, 905, 604, 702, 710

Fatralyses 1 193vasoi Road, Boxb Don A 01 NOTetra.c SH0T57P72

SS0513-2 Rev. 02/18