



# Calibration Certificate

Technician: SO *SO*  
Part No: 2781500MA1B2BT1  
Model: 278

Serial No: 4999333      Range: 512 to 1100 HPA/MB  
Work Order: 24021876      Nom. Output: 0.1 to 5 VDC  
Date: 02/14/2012      Supply: 24vdc

---

## CALIBRATION DATA

---

APPLIED PRESSURE (hPa)	TRANSDUCER OUTPUT (VDC)	PRESSURE CONVERSION (hPa)	ERROR (hPa)	EQUIPMENT UNCERTAINTY (hPa)
511.99	0.0983	511.80	-0.19	+/- 0.10
658.99	1.3272	659.26	0.27	+/- 0.10
806.00	2.5514	806.17	0.17	+/- 0.10
952.99	3.7740	952.88	-0.11	+/- 0.10
1100.01	5.0014	1100.17	0.16	+/- 0.10

---

### AMBIENT CONDITIONS:

Humidity: 15.0 %RH  
Pressure: 1003.7 hPa

Temperature: 26.7 degree C

### SPECIFICATIONS:

Accuracy Specification: +/- 0.6 hPa @ +20 degree C (+68 degree F).

### NOTES:

1. This calibration was performed in compliance with ANSI/NC SL Z540-1-1994.
2. All errors are expressed in hPa.
3. Consult specification sheet for additional information.
4. This calibration is certified per N.I.S.T. traceable primary standards.  
Reference standard: I/N\_00049-SN237-DHI PPC3-2M A1.4MS/A160KP.  
Reference standard cal. due date: 10/26/11 - 10/26/12.
5. This certificate may not be reproduced, except in full, without written approval from Setra Systems.
6. This calibration was performed using procedure P2781X.

159 Swanson Road, Boxborough, MA 01719/Telephone 1-800-257-3872, (978) 263-1400