



Unser Zeichen
Hallberg Katja
Spediteur/Paketdienst
UPS Courier

SERVICE REPORT

Datum
26.10.2012

Referenz des Verkäufers
106238

Referenz des Käufers
0030018219

Page 1

Auftraggeber
Stiftung Alfred-Wegener-Institut für Polar- und
Meeresforschung
Geschäftszimmer/Gebäude E
Am Handelshafen 12
27570 Bremerhaven
Germany

**Rechnungsadresse (wenn nicht
Auftraggeber)**

Pos	Beschreibung	Seriennummer -	Menge
103180	HMT330 Humidity and Temperature Transmitter --CUSTOMER'S DESCRIPTION OF FAILURE-- Repair and calibration. --PROBLEM(S) FOUND -- Probe has a bad connection, no stable humidity results. Old sensor was out of tolerance and there was some fault on errorlog. --ACTION(S) TAKEN-- Installed new probe and sensor. Installed filter for probe, it was missing. Operation tested, adjustments made and unit was calibrated. Calibration certificate numbers H37-12430010 (analog output) and H35-12430024 (after adjustment) issued.	A4650017	1
103182	HMT330 Humidity and Temperature Transmitter --CUSTOMER'S DESCRIPTION OF FAILURE-- Calibration. --PROBLEM(S) FOUND -- No fault found. --ACTION(S) TAKEN-- Installed filter for probe, it was missing. Operation tested, adjustments made and unit was calibrated. Calibration certificate numbers H37-12430001 (analog output), H35-12420046 (before adjustment) and H35-12420053 (after adjustment) issued.	A4650018	1

Vaisala Oyj
Vanha Nurmijärventie 21
FI-01670, Vantaa
P.O. Box 26, FI-00421 Helsinki
Finland

Telefon +358 9 8949 1
Telefax +358 9 8949 2227

Sitz Vantaa, Finland

Bankverbindung
Nordea Bank Finland Oyj
FI4815713000014119
BIC/SWIFT NDEAFIHH

Ust. - Nr. FI01244162
FK-Nummer 0124416-2