



Unser Zeichen
Jördens Doris (Doris)

SERVICE REPORT

Datum
07-JUN-2018

SR#
1222253

Referenz des Käufers
39/0045116123/NB

Auftraggeber
Alfred-Wegener-Institut
Flughafen Bremen
AWI-Hangar E
Heinrich-Focke-Str. 9
28199 Bremen
Germany

Rechnungsadresse (wenn nicht Auftraggeber)
Alfred-Wegener-Institut
Helmholtz-Zentrum für Polar- und Meeresforschung
Telegrafenberg A45
14473 Potsdam
Germany

Service Order	Beschreibung	Seriennummer	Menge								
508895	<p>HMT330 Premium Humidity and Temperature Transmitter with metal enclosure</p> <p>--REASON FOR RETURN-- Repair and Standard humidity calibration. 5 points, 15, 33, 54, 75, 95% RH at ambient temperature (RHCALSTD).</p> <p>Customer wishes rebuild of sensor and change from HUMICAP180C to HUMICAP180R</p> <p>--PROBLEM(S) FOUND -- Cable modified by customer.</p> <p>--ACTION(S) TAKEN-- Sensor changed from HUMICAP180C to HUMICAP180R. Operation checked, adjustments made.</p> <table border="0"> <tr> <td>Reference humidity</td> <td>Observed humidity</td> </tr> <tr> <td>15.0%RH</td> <td>14.5%RH</td> </tr> <tr> <td>49.9%RH</td> <td>49.4%RH</td> </tr> <tr> <td>94.9%RH</td> <td>95.2%RH</td> </tr> </table>	Reference humidity	Observed humidity	15.0%RH	14.5%RH	49.9%RH	49.4%RH	94.9%RH	95.2%RH	M4650074	1
Reference humidity	Observed humidity										
15.0%RH	14.5%RH										
49.9%RH	49.4%RH										
94.9%RH	95.2%RH										