

# CALIBRATION CERTIFICATE

**Instrument** Humidity and Temperature Transmitter HMT333  
**Order code** HMT330 3F0A002BCAC100A04AABAA1  
**Serial number** A4650017  
**Manufacturer** Vaisala Oyj, Finland  
**Calibration date** 29th January 2008  
**Test procedure** doc210426-a

The above instrument was calibrated by comparing the readings of the instrument to working standards of the manufacturer. The reference humidity was calculated from dewpoint temperature and temperature readings with the exception of the driest condition that was measured as relative humidity. Dewpoint temperature was measured with a 373 LHX dewpoint meter. Temperature and relative humidity were measured with two factory working standards. At the time of shipment, the instrument described above met its operating specifications.

The 373 LHX dewpoint meter has been calibrated at Vaisala Measurement Standards Laboratory (MSL) by using a MSL working standard traceable to National Institute of Standards and Technology (NIST). The temperature readings of the factory working standards have been calibrated at MSL by using MSL working standards traceable to NIST. The relative humidity readings of the factory working standards have been calibrated at the Vaisala factory by using a 373 LHX dewpoint meter. The temperature calibration at MSL has been accredited by the FINAS according to the ISO/IEC 17025.

## Humidity calibration results

Reference humidity %RH	Reference temperature °C	Observed humidity %RH	Observed probe temperature °C	Additional probe temperature °C	Humidity difference %RH	Permissible difference %RH
+ 93.5	+ 22.19	+ 93.5	+ 22.19	-	0.0	± 1.7
+ 73.9	+ 22.18	+ 74.0	+ 22.17	-	+ 0.1	± 1.0
+ 53.4	+ 22.16	+ 53.5	+ 22.16	-	+ 0.1	± 1.0
+ 32.7	+ 22.17	+ 33.0	+ 22.16	-	+ 0.3	± 1.0
+ 12.4	+ 22.18	+ 12.2	+ 22.18	-	- 0.2	± 1.0
+ 0.1	+ 22.19	- 0.2	+ 22.18	-	- 0.3	± 1.0

## Temperature calibration results

Reference temperature °C	Observed probe temperature °C	Temperature difference °C	Additional probe temperature °C	Temperature difference °C	Permissible difference °C
+ 22.18	+ 22.17	- 0.01	-	-	± 0.10

## Equipment used in calibration

Type	Serial number	Calibration date	Certificate number
373 LHX	06-0122	2007-08-03	M-07H053
Vaisala HMP233 / T	623073	2008-01-25	K008-R00190
Vaisala HMP233 / T	P1740020	2008-01-25	K008-R00191
HMP233 / RH	623073	2008-01-28	H33-08051001
HMP233 / RH	P1740020	2008-01-28	H33-08051002

## Uncertainties ( 95 % confidence level, k=2)

Humidity ± 0.6%RH @ 0...40%RH, ± 1.0%RH @ 40...97%RH

Temperature ± 0.10 °C.

**Ambient conditions** / Humidity 33 ± 5%RH, Temperature 23 ± 1 °C, Pressure 1007 ± 1 hPa.

For Vaisala Oyj

Sari Kaitanen

*This report shall not be reproduced except in full, without the written approval of Vaisala.*

doc212778-b