



CERTIFICATE OF CONFORMANCE

Customer Name	Model #	Serial #	Date
Alfred Wegner Institute	CR-2	210	8/05/2009

NIST TRACEABLE STANDARD	CUSTOMER UNIT DEW POINT	DIFFERENCE	SPECIFICATION OF UNIT
10.37°C	10.41°C	+0.04°C	±0.10°C
1.19°C	1.21°C	+0.02°C	±0.10°C
-20.91°C	-20.90°C	+0.01°C	±0.10°C
-30.52°C	-30.50°C	+0.02°C	±0.10°C
-72.9°C	-72.90°C	-0.0°C	±0.10°C

Ambient Temp: 72 °F

Relative Humidity: 50%RH

NIST TRACEABLE CALIBRATION EQUIPMENT

Humidity Standard

Manufacturer	Model	Serial #
Edgetech	DewPrime 1	26073/3C975M
Buck Research Instruments	CR-3 Cryogenic Hygrometer	301

Temperature Standard

Manufacturer	Model	Serial #
Logan	PlatinumRTD	9251-12

Voltage Standard

Manufacturer	Model	Serial #
Agilent	34401A	US36112687
Hewlett Packard	34401A	US36043947

Pressure Standard

Manufacturer	Model	Serial #
Druck	DPI-740	740 01260

TESTED BY: JLP/BA

PO Box 19498 Boulder, CO 80308
main: 303.442.6055
fax: 303.443.2986
1.800.HUMIDITY
WWW.HYGROMETERS.COM