

ROSEMOUNT AEROSPACE INC.  
PLATINUM RESISTANCE THERMOMETER

REPORT OF CALIBRATION  
IPTS-1968

MODEL: 0102EK1BB                      SERIAL: B10682                      05-04-2012  
PART : 00102-1193-0001  
ATP : D9630048    REV: B

THIS TABLE WAS COMPUTER CHECKED FOR INTERCHANGEABILITY AGAINST A MASTER TABLE.  
AN '\*' ADJACENT TO RESISTANCE INDICATES AN OUT OF TOLERANCE CONDITION.

CALIBRATION CONSTANTS & TABLE

ELEMENT # 1

-----  
R0        =    99.96362  
R100      =    139.33714  
ALPHA     =    0.00393879  
DELTA     =    1.49170357  
BETA      =    0.11000000

TEMP deg C	RESISTANCE ohms
-----	-----
0.000	99.964
100.000	139.337