



Kalibrierprotokoll		
Sensor Rosemount 1221VL6B1B	Serien-Nr. 1013	
Druckreferenz MPS 30	Serien-Nr. 5029	Messbereich +/-17,2 hPa
Spannungsmessgerät Keithley 2000	Serien-Nr. 703709	T _{Umgeb.} [°C] 21
Dateiname Keithley_20090106_diff.adf	Kanal 0	

Referenz [hPa]	Spannung [V]	Berechnet [hPa]	Abweichung [hPa]
17	9,9	17,0	-0,01
15	8,74	15,0	0,00
13	7,572	13,0	0,00
11	6,415	11,0	0,01
9	5,25	9,0	0,01
7	4,08	7,0	0,00
5	2,913	5,0	0,00
0	-0,003	0,0	0,00
-5	-2,92	-5,0	-0,01
-7	-4,085	-7,0	-0,01
-9	-5,25	-9,0	-0,01
-11	-6,41	-11,0	0,00
-13	-7,57	-13,0	0,01
-15	-8,734	-15,0	0,01
-17	-9,905	-17,0	0,00

Werte der Ausgleichsgeraden

Offset 0,000800928 hPa
Steigung 1,716275042 hPa/V



06.01.2009

Datum

Name