

| | | |
|---|---|---|
| <i>Rev:</i> 1.00 <i>Date:</i> 24.01.2014 <i>Page:</i> 1 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| <i>Document-P/N:</i> | CDR-ADP121103 | |

AVENTECH RESEARCH INC.

MODEL AIMMS20-ADP

CALIBRATION DATA REPORT

S/N ADP121103

Document-P/N: *CDR-ADP121103*


**Aventech Research Inc.
110 Anne Street South, Unit 23
Barrie, Ontario, Canada L4N 2E3**

**Tel: (705) 722-4288
Fax: (705) 722-9077**

Web: www.aventech.com

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 2 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

Issued

| Date | Name | Signature | Function |
|------------|--------------|--|----------|
| 24.01.2014 | Colin Speake |  | PM |

Checked

| Date | Name | Signature | Function |
|------------|----------------|--|----------|
| 24.01.2014 | Stephen Foster |  | DE |

Approved

| Date | Name | Signature | Function |
|------------|----------------|--|----------|
| 24.01.2014 | Stephen Foster |  | DE |

Customer Approval (as required)

| Date | Name | Signature | Function |
|------|------|-----------|----------|
| | | | |

Table of Functions

| | |
|-------------|-----------------------------------|
| CUST | Customer |
| DE | Design Engineer |
| HD | Head of Design Engineering |
| PM | Production Manager |
| QM | Quality Manager |

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 3 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

RESTRICTED DISCLOSURE NOTICE

The information, technical specifications and technical data contained within this document are proprietary and confidential to Aventech Research Inc. and constitute trade secrets for purposes of the Trade Secret and Freedom of Information Acts. The contents of this document shall not be used for purposes other than for which it is released. No disclosures to others, either in Canada, the United States or abroad, or reproduction of any part of the information released is to be made, except by written authorization of Aventech Research Inc.

© 2011 Aventech Research Inc., Canada.
All Rights Reserved.

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 4 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

RECORD OF REVISIONS

| Date | Issue | Page | Paragraph | Firmware Revision | Comments |
|------------|-------|------|-----------|-------------------|----------------|
| 24.01.2014 | 1.00 | All | All | | Final, Issue 1 |

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 5 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

1 Contents

| | | |
|-----|-----------------------------------|----|
| 2 | OVERVIEW | 6 |
| 3 | SENSOR CALIBRATION DATA SET | 7 |
| 3.1 | Barometric Sensor | 7 |
| 3.2 | Angle-of-Attack | 7 |
| 3.3 | Angle-of-Sideslip | 8 |
| 3.4 | Pitot-Static | 8 |
| 3.5 | Temperature Sensor | 9 |
| 3.6 | Relative Humidity Sensor..... | 9 |
| 3.7 | Acceleration Sensors..... | 10 |
| 3.8 | Geomagnetic Field Sensors | 10 |

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 6 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

PROBE SERIAL NUMBER: ADP121103

2 OVERVIEW

The air-data probe (ADP) data system is comprised of eight sensor subsystems listed below:

1. Barometric pressure
2. Angle-of-attack differential pressure
3. Sideslip differential pressure
4. Pitot-static differential pressure
5. Temperature sensor (thermistor)
6. Relative humidity sensor
7. Acceleration sensors (x 3 channels)
8. Geomagnetic field sensors (x 3 channels)

The ADP digital signal processor (DSP) reads the digitized signals from the analog-to-digital converter (ADC) and applies various calibration equations to derive fully compensated and calibrated sensor data. The parameters associated with these calibration equations are detailed in Section 3 for each sensor subsystem.

Calibration parameters are stored in non-volatile DFLASH memory on-board the DSP and can be reprogrammed through an RS232 serial interface.

Pressure sensors are temperature compensated using a special ASIC device, one dedicated to each sensor. Analog signals from the pressure sensors are compensated before they are input to the ADC. Therefore, temperature compensation is not performed by the DSP for these sensors and temperature compensation parameters are not included in the data set stored by the internal DFLASH unit.

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 7 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

3 SENSOR CALIBRATION DATA SET

Sensor calibration functions are expressed in terms of a normalized channel voltage, i.e. voltage expressed as a fraction of full scale. Thus, with a 4.096V reference, the voltage for each channel is expressed as

$$\tilde{V} = \frac{V}{4.096}$$

3.1 Barometric Sensor

Calibration Function: $P = C_0 + C_1\tilde{V}$
Units: Pa

As of Oct. 31, 2011:

| Coefficient | Value |
|-------------|-----------|
| C_0 | 5666.41 |
| C_1 | 110014.80 |

As of May. 7, 2013:

| Coefficient | Value |
|-------------|-----------|
| C_0 | 5757.75 |
| C_1 | 110011.40 |

3.2 Angle-of-Attack

Calibration Function: $P = C_0 + C_1\tilde{V}$
Units: Pa

As of Oct. 31, 2011:

| Coefficient | Value |
|-------------|----------|
| C_0 | -6998.43 |
| C_1 | 14018.84 |

As of May. 7, 2013:

| Coefficient | Value |
|-------------|----------|
| C_0 | -7002.35 |
| C_1 | 14019.07 |

3.3 Angle-of-Sideslip

Calibration Function: $P = C_0 + C_1 \tilde{V}$
Units: Pa

As of Oct. 31, 2011:

| Coefficient | Value |
|-------------|----------|
| C_0 | -7005.71 |
| C_1 | 14018.76 |

As of May. 7, 2013:

| Coefficient | Value |
|-------------|----------|
| C_0 | -7008.48 |
| C_1 | 14015.85 |

3.4 Pitot-Static

Calibration Function: $P = C_0 + C_1 \tilde{V}$
Units: Pa

As of Oct. 31, 2011:

| Coefficient | Value |
|-------------|----------|
| C_0 | -1281.20 |
| C_1 | 14007.77 |

As of May. 7, 2013:

| Coefficient | Value |
|-------------|----------|
| C_0 | -1289.62 |
| C_1 | 14020.81 |

| | | |
|--|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 9 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

3.5 Temperature Sensor

Calibration Function: $T = C_0 + C_1\tilde{V} + C_2\tilde{V}^2 + C_3\tilde{V}^3$
Units: C

As of Oct. 31, 2011:

| Coefficient | Value – Low Range (-40C to -15C) | Value – High Range (-15C to +50C) |
|-------------|-------------------------------------|--------------------------------------|
| C_0 | 6341.965 | 159.9910 |
| C_1 | -27423.72 | -629.8731 |
| C_2 | 39696.26 | 981.5250 |
| C_3 | -19291.20 | -641.7212 |

As of May. 7, 2013:

| Coefficient | Value – Low Range (-40C to -15C) | Value – High Range (-15C to +50C) |
|-------------|-------------------------------------|--------------------------------------|
| C_0 | 69341.665 | 159.691 |
| C_1 | -27423.72 | -629.8731 |
| C_2 | 39696.26 | 981.5250 |
| C_3 | -19291.2 | -641.7212 |

3.6 Relative Humidity Sensor

Calibration Function: $RH = C_0 + C_1\tilde{V}$
Units: fractional value, 0 - 1

As of Oct. 31, 2011:

| Coefficient | Value |
|-------------|-----------|
| C_0 | 1.0959 |
| C_1 | -1.404873 |

As of May. 7, 2013:

| Coefficient | Value |
|-------------|---------|
| C_0 | 1.0959 |
| C_1 | -1.4049 |

| | | |
|---|---|---|
| Rev: 1.00 Date: 24.01.2014 Page: 10 of 10 | AVENTECH RESEARCH INC. MODEL AIMMS20-ADP CALIBRATION DATA REPORT |  |
| Document-P/N: | CDR-ADP121103 | |

3.7 Acceleration Sensors

Calibration Function: $A_{x,y,z} = (C_0 + C_1 \tilde{V}_t) + (C_2 + C_3 \tilde{V}_t) \tilde{V}_{x,y,z}$

Units: m/s/s, \tilde{V}_t = sensor-temperature signal

As of Oct. 31, 2011:

| Coefficient | Value - x | Value - y | Value - z |
|-------------|------------|------------|------------|
| C_0 | -95.820620 | -88.307940 | -77.269970 |
| C_1 | -0.532949 | -3.508326 | 1.651912 |
| C_2 | 217.7026 | 222.7078 | 228.7866 |
| C_3 | -0.998233 | 1.749547 | 8.139198 |

As of May. 7, 2013:

| Coefficient | Value - x | Value - y | Value - z |
|-------------|-----------|-----------|-----------|
| C_0 | -86.0209 | -84.5728 | -91.1055 |
| C_1 | -6.8434 | -8.8828 | -12.3439 |
| C_2 | 214.9938 | 218.1814 | 214.9037 |
| C_3 | 4.6590 | 3.2845 | 6.2206 |

3.8 Geomagnetic Field Sensors

Calibration Function: $M_{x,y,z} = C_0 + C_1 \tilde{V}_{x,y,z}$

Units: μT

As of Oct. 31, 2011:

| Coefficient | Value-x | Value-y | Value-z |
|-------------|---------|---------|---------|
| C_0 | -110.00 | 109.23 | -111.34 |
| C_1 | 220.00 | -201.71 | 213.18 |

As of May. 7, 2013:

| Coefficient | Value-x | Value-y | Value-z |
|-------------|---------|---------|---------|
| C_0 | -110.00 | 109.23 | -111.34 |
| C_1 | 220.00 | -201.71 | 213.18 |