

CERTIFICATE OF CONFORMANCE

Customer Name	Model #	Serial #	Date
AWI	CR-2	218	10/15/2011

NIST TRACEABLE STANDARD	CUSTOMER UNIT DEW POINT	DIFFERENCE	SPECIFICATION OF UNIT
4.96°C	4.99°C	+0.03°C	±0.10°C
-11.35°C	-11.29°C	+0.06°C	±0.10°C
-31.70°C	-31.67°C	+0.03°C	±0.10°C
-65.10°C	-65.09°C	+0.01°C	±0.10°C

Ambient Temp: 72 °F**Relative Humidity: 50%RH****NIST TRACEABLE CALIBRATION EQUIPMENT****Humidity Standard**

Manufacturer	Model	Serial #
Edgetech	DewPrime 1	26073/3C975M
Buck Research Instruments	CR-3 Cryogenic Hygrometer	301

Temperature Standard

Manufacturer	Model	Serial #
Logan	PlatinumRTD	9251-12

Voltage Standard

Manufacturer	Model	Serial #
Agilent	34401A	US36112687
Hewlett Packard	34401A	US36043947

Pressure Standard

Manufacturer	Model	Serial #
Druck	DPI-740	740 01260

TESTED BY: 