

## Calibration protocol

pressure sensor [V] to [hPa]

date: 13.11.2012

data file:

AWI-P-Cal\_20121113\_2.mat

sensor:

type: Rosemount 1221F2VL3B1B

sensor no.: S/N 1013

reference pressure sensor:

type: Weston Aerospace DPM7885

ref. sensor no.: S/N 150227 and 150226

reference voltage sensor:

type: Keithley 2000 ref. sensor no.: S/N 703709

coefficients to compute SN1013:

[0.000008, 0.000074, 1.721668, -0.026]

coefficients to compute 1st order from SN1013:

[1.722285, -0.023]

