

## Calibration protocol

pressure sensor [V] to [hPa]

date: 13.11.2012

data file:

AWI-P-Cal\_20121113\_4.mat

sensor:

type: Rosemount 1221F2VL2B1B

sensor no.: S/N 557

reference pressure sensor:

type: Weston Aerospace DPM7885

ref. sensor no.: S/N 150227 and 150226

reference voltage sensor:

type: Keithley 2000 ref. sensor no.: S/N 703709

coefficients to compute SN557:

[-0.000034, 0.000275, 5.154337, 0.027]

coefficients to compute 1st order from SN557:

[5.152276, 0.036]

