

## Calibration protocoll

pressure sensor [V] to [hPa]

date: 13.11.2012

data file:

AWI-P-Cal\_20121113\_4.mat

sensor:

type: Rosemount 1221F2VL6B1B

sensor no.: S/N 487

reference pressure sensor:

type: Weston Aerospace DPM7885

ref. sensor no.: S/N 150227 and 150226

reference voltage sensor:

type: Keithley 2000 ref. sensor no.: S/N 703709

coefficients to compute SN487:

[-0.000018,0.000445,5.162359,0.007]

coefficients to compute 1st order from SN487:

[5.161420,0.023]

